
Midmark M11 - v1.0.5
Total Cycles: 3333

Name: RITTER2
Sterilizer ID: V2313228
01/29/2024 04:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.3 F 0.0
2:00 187.6 F 3.6
4:00 213.2 F 9.6
6:00 224.6 F 6.5
8:00 240.2 F 12.7
10:00 240.5 F 11.2
12:00 240.3 F 10.7
14:00 255.8 F 18.6
16:00 268.9 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 271.8 F 28.8
2:00 271.9 F 28.9
3:00 271.9 F 28.9
4:00 271.7 F 28.6

Min 271.7 F 28.4
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 218.9 F 0.0
Dry Cycle Aborted

FILLING: 2:31
HEATING: 16:54
STERILIZING: 4:00
VENTING: 2:51
DRYING: 0:13
TOTAL CYCLE: 00:26:29

01/29/2024 04:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____