
Midmark M11 - v1.0.5
Total Cycles: 678

Name: RITTER2
Sterilizer ID: V2313228
01/10/2022 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.6 F 0.0
2:00 157.3 F 2.8
4:00 175.0 F 6.6
6:00 196.6 F 11.6
8:00 217.1 F 6.3
10:00 229.8 F 11.7
12:00 233.8 F 9.3
14:00 237.4 F 9.6
16:00 247.9 F 14.8
18:00 260.1 F 21.6
20:00 271.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.6
1:00 272.9 F 30.1
2:00 271.8 F 29.5
3:00 273.0 F 30.3
4:00 272.3 F 29.7

Min 271.6 F 29.4
Max 273.3 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 216.8 F 0.0
10:00 226.6 F 0.0
15:00 224.3 F 0.0
20:00 224.0 F 0.0
25:00 222.9 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:40
CYCLE COMPLETE

01/10/2022 04:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____