

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3652

Name: RITTER 1  
Sterilizer ID: V2178198  
08/15/2023 09:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.3 F 0.0  
2:00 186.6 F 2.4  
4:00 191.8 F 6.0  
6:00 205.1 F 11.1  
8:00 219.3 F 5.5  
10:00 231.7 F 10.4  
12:00 232.0 F 7.8  
14:00 243.0 F 12.8  
16:00 241.1 F 11.1  
18:00 253.8 F 17.5  
20:00 264.8 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.0 F 28.6  
2:00 272.3 F 29.1  
3:00 271.8 F 29.1  
4:00 271.6 F 28.4

Min 271.4 F 28.3  
Max 272.4 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.9 F 0.0  
5:00 233.0 F 0.0  
10:00 228.5 F 0.0  
15:00 228.3 F 0.0  
20:00 228.5 F 0.0  
25:00 228.8 F 0.0  
30:00 228.5 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 21:52  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:00:48  
CYCLE COMPLETE

08/15/2023 10:12 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_