

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4230

Name: RITTER2  
Sterilizer ID: V2313228  
12/19/2024 11:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.6 F 0.0  
2:00 131.5 F 0.5  
4:00 134.0 F 2.8  
6:00 163.4 F 6.1  
8:00 185.8 F 9.9  
10:00 201.0 F 8.6  
12:00 217.1 F 6.5  
14:00 229.9 F 10.9  
16:00 229.9 F 7.4  
18:00 241.7 F 12.9  
20:00 241.3 F 11.4  
22:00 253.2 F 17.5  
24:00 264.0 F 24.0  
26:00 272.4 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 271.9 F 29.7  
2:00 272.9 F 29.9  
3:00 272.1 F 29.5  
4:00 271.6 F 29.1

Min 271.1 F 28.8  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.7 F 0.0  
5:00 216.5 F 0.0  
10:00 229.8 F 0.0  
15:00 221.4 F 0.0  
20:00 228.8 F 0.0  
25:00 230.3 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 26:16  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:05:09  
CYCLE COMPLETE

12/19/2024 12:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_