

-----  
Midmark M11 - v1.0.5  
Total Cycles: 972

Name: RITTER2  
Sterilizer ID: V2313228  
04/11/2022 08:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.0 F 0.0  
2:00 175.8 F 0.1  
4:00 174.8 F 0.6  
6:00 201.3 F 4.7  
8:00 216.8 F 9.9  
10:00 227.2 F 8.0  
12:00 236.4 F 11.6  
14:00 237.0 F 9.9  
16:00 237.3 F 9.0  
18:00 250.6 F 15.9  
20:00 262.6 F 22.9  
22:00 272.6 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.5 F 29.5  
2:00 271.6 F 28.7  
3:00 272.1 F 29.6  
4:00 272.4 F 29.2

Min 271.3 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.1 F 0.0  
5:00 208.3 F 0.0  
10:00 220.9 F 0.0  
15:00 217.6 F 0.0  
20:00 213.9 F 0.0  
25:00 213.1 F 0.0  
30:00 222.3 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 22:17  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:00:26  
CYCLE COMPLETE

04/11/2022 09:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_