
Midmark M11 - v1.0.0
Total Cycles: 1419

Name: RITTER 1
Sterilizer ID: V2178198
12/01/2021 05:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.7 F 0.0
2:00 144.0 F 1.2
4:00 174.1 F 4.1
6:00 202.0 F 8.6
8:00 217.2 F 12.4
10:00 225.3 F 7.2
12:00 237.0 F 12.2
14:00 234.8 F 9.0
16:00 239.5 F 10.0
18:00 246.1 F 13.5
20:00 258.3 F 20.0
22:00 269.3 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.5
1:00 271.7 F 28.8
2:00 272.8 F 29.2
3:00 272.0 F 28.7
4:00 271.5 F 28.6

Min 271.4 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 202.4 F 0.0
10:00 203.7 F 0.0
15:00 210.4 F 0.0
20:00 214.8 F 0.0
25:00 216.8 F 0.0
30:00 217.7 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 22:56
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:49
CYCLE COMPLETE

12/02/2021 10:46 AM
ACCEPTED: X
REJECTED:
Operator: U02