
Midmark M11 - v1.0.5
Total Cycles: 3838

Name: RITTER2
Sterilizer ID: V2313228
08/08/2024 08:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 72.6 F 0.0
4:00 82.1 F 0.0
6:00 103.3 F 0.0
8:00 143.5 F 0.8
10:00 171.3 F 3.5
12:00 189.0 F 6.6
14:00 206.8 F 10.1
16:00 219.9 F 7.8
18:00 225.8 F 7.8
20:00 235.7 F 12.0
22:00 233.8 F 8.7
24:00 242.4 F 12.0
26:00 243.8 F 12.4
28:00 255.1 F 18.3
30:00 265.2 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.0 F 28.8
2:00 272.4 F 29.3
3:00 271.8 F 29.2
4:00 271.7 F 28.7

Min 271.4 F 28.6
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 224.0 F 0.0
10:00 200.6 F 0.0
15:00 195.1 F 0.0
20:00 191.6 F 0.0
25:00 188.5 F 0.0
30:00 192.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:53
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:10:05
CYCLE COMPLETE

08/08/2024 09:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____