
Midmark M11 - v1.0.0
Total Cycles: 4695

Name: RITTER 1
Sterilizer ID: V2178198
07/03/2024 09:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.2 F 0.0
2:00 171.0 F 2.5
4:00 171.6 F 6.0
6:00 188.2 F 9.8
8:00 210.2 F 7.5
10:00 225.7 F 7.6
12:00 228.1 F 12.3
14:00 233.2 F 8.6
16:00 240.5 F 11.2
18:00 244.0 F 12.4
20:00 255.7 F 18.5
22:00 266.3 F 25.0
24:00 272.5 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.0 F 28.9
2:00 272.8 F 29.2
3:00 272.3 F 29.4
4:00 271.6 F 29.1

Min 271.5 F 28.8
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.8 F 0.0
5:00 239.2 F 0.0
10:00 225.3 F 0.0
15:00 220.8 F 0.0
20:00 207.1 F 0.0
25:00 209.2 F 0.0
30:00 210.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 24:24
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:02:60
CYCLE COMPLETE

07/03/2024 10:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____