
Midmark M11 - v1.0.0
Total Cycles: 1714

Name: RITTER 1
Sterilizer ID: V2178198
03/14/2022 12:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.0 F 0.0
2:00 186.8 F 3.9
4:00 202.2 F 8.9
6:00 219.1 F 7.8
8:00 229.4 F 9.1
10:00 236.8 F 9.8
12:00 242.9 F 12.4
14:00 242.1 F 11.2
16:00 255.3 F 18.0
18:00 266.9 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.7 F 29.0
2:00 271.9 F 28.3
3:00 271.8 F 28.5
4:00 272.7 F 29.0

Min 271.4 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 240.0 F 0.0
10:00 227.6 F 0.0
15:00 229.0 F 0.0
20:00 225.8 F 0.0
25:00 226.6 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:15
CYCLE COMPLETE

03/14/2022 01:52 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____