
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
2.5uS / 1.6ppm
CYCLE NUMBER 000045
5:03 25/08/2020
WRAPPED
134 C for 6 min.
CYCLE START 0:00
CONDITIONING 6:24
PRESSURIZING 18:47
STERILIZING 21:05
136.3 C 226kPa 21:05
136.4 C 227kPa 21:35
136.4 C 227kPa 22:05
136.4 C 224kPa 22:35
136.5 C 227kPa 23:05
136.2 C 225kPa 23:35
136.4 C 226kPa 24:05
136.3 C 226kPa 24:35
136.3 C 227kPa 25:05
136.3 C 227kPa 25:35
136.4 C 227kPa 26:05
136.4 C 227kPa 26:35
136.3 C 227kPa 27:05
VENTING 27:05
AIR DRYING 28:16
CYCLE COMPLETE 58:16

Digital Signature #
83BE3B8078B49CEA
