
Midmark M11 - v1.0.0
Total Cycles: 3182

Name: RITTER 1
Sterilizer ID: V2178198
05/01/2023 12:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.2 F 0.0
2:00 190.4 F 3.7
4:00 193.3 F 8.2
6:00 211.0 F 13.4
8:00 224.2 F 8.3
10:00 235.3 F 11.5
12:00 240.0 F 11.8
14:00 239.7 F 10.7
16:00 253.8 F 17.8
18:00 265.9 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.3 F 29.3
2:00 271.6 F 29.1
3:00 272.3 F 29.3
4:00 271.4 F 28.9

Min 271.4 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 210.7 F 0.0
10:00 228.8 F 0.0
15:00 237.6 F 0.0
20:00 233.3 F 0.0
25:00 225.3 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 19:34
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

05/01/2023 01:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____