
Midmark M11 - v1.0.5
Total Cycles: 2950

Name: RITTER2
Sterilizer ID: V2313228
09/29/2023 11:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.4 F 0.0
2:00 172.5 F 2.7
4:00 186.8 F 6.8
6:00 206.4 F 11.4
8:00 218.7 F 6.1
10:00 229.3 F 11.0
12:00 232.2 F 7.8
14:00 242.5 F 12.4
16:00 244.3 F 12.5
18:00 256.1 F 18.8
20:00 266.6 F 25.3
22:00 273.2 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 271.6 F 29.2
2:00 271.9 F 29.3
3:00 272.5 F 29.5
4:00 272.8 F 29.7

Min 271.3 F 29.2
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.0 F 0.0
5:00 232.7 F 0.0
10:00 214.2 F 0.0
15:00 188.3 F 0.0
20:00 183.0 F 0.0
25:00 182.3 F 0.0
30:00 189.3 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 22:16
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:01:27
CYCLE COMPLETE

09/29/2023 12:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____