
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000267
7:34 01/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:43
STERILIZING 5:20
136.0 C 222kPa 5:20
136.2 C 232kPa 5:50
136.3 C 232kPa 6:20
136.3 C 231kPa 6:50
136.4 C 233kPa 7:20
136.5 C 232kPa 7:50
136.3 C 232kPa 8:20
136.3 C 232kPa 8:50
136.3 C 232kPa 8:50
VENTING 8:51
AIR DRYING 9:50
CYCLE COMPLETE 24:50

Digital Signature #
330705F09DA280E0
