

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4684

Name: RITTER 1  
Sterilizer ID: V2178198  
06/27/2024 11:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.1 F 0.0  
2:00 119.1 F 1.0  
4:00 125.5 F 3.2  
6:00 150.0 F 6.1  
8:00 171.7 F 9.4  
10:00 202.9 F 13.3  
12:00 212.6 F 5.9  
14:00 227.1 F 9.8  
16:00 233.2 F 9.9  
18:00 236.8 F 10.5  
20:00 234.8 F 8.1  
22:00 246.5 F 13.8  
24:00 257.4 F 19.7  
26:00 267.4 F 26.1  
28:00 273.1 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.3  
1:00 272.9 F 29.6  
2:00 272.2 F 29.8  
3:00 272.0 F 29.3  
4:00 272.7 F 29.9

Min 271.5 F 29.2  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.8 F 0.0  
5:00 213.9 F 0.0  
10:00 222.5 F 0.0  
15:00 235.7 F 0.0  
20:00 229.2 F 0.0  
25:00 226.6 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 28:17  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:06:53  
CYCLE COMPLETE

06/27/2024 12:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_