
Midmark M11 - v1.0.5
Total Cycles: 1900

Name: RITTER2
Sterilizer ID: V2313228
01/05/2023 12:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.2 F 0.0
2:00 166.2 F 3.3
4:00 181.8 F 7.2
6:00 202.5 F 12.1
8:00 217.1 F 6.2
10:00 229.7 F 11.3
12:00 231.5 F 8.3
14:00 240.9 F 11.4
16:00 244.3 F 13.1
18:00 256.7 F 19.6
20:00 267.7 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.5 F 29.8
2:00 271.6 F 29.1
3:00 272.5 F 30.0
4:00 272.1 F 29.4

Min 271.5 F 29.1
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 234.2 F 0.0
10:00 235.1 F 0.0
15:00 216.8 F 0.0
20:00 210.5 F 0.0
25:00 213.7 F 0.0
30:00 215.5 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 21:23
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:48
CYCLE COMPLETE

01/05/2023 01:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____