
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 003766
13:15 06/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:08
STERILIZING 4:11
136.5 C 220kPa 4:11
136.4 C 223kPa 4:41
136.0 C 221kPa 5:11
136.3 C 224kPa 5:41
136.3 C 224kPa 6:11
136.4 C 224kPa 6:41
136.3 C 222kPa 7:11
136.4 C 225kPa 7:41
136.4 C 225kPa 7:41
VENTING 7:41
AIR DRYING 8:30
CYCLE COMPLETE 8:50

Digital Signature #
343F93F6D23FCC1F
