
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002648
8:20 03/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:35
STERILIZING 3:39
136.0 C 220kPa 3:39
136.3 C 224kPa 4:09
136.4 C 225kPa 4:39
136.4 C 222kPa 5:09
136.4 C 224kPa 5:39
136.1 C 223kPa 6:09
136.4 C 225kPa 6:39
136.4 C 222kPa 7:09
136.4 C 222kPa 7:09
VENTING 7:09
AIR DRYING 8:01
CYCLE COMPLETE 9:12

Digital Signature #
1F3008C848533E7E
