

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000828  
15:20 17/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:31  
PRESSURIZING 2:38  
STERILIZING 3:39  
136.0 C 220kPa 3:39  
136.2 C 231kPa 4:09  
136.3 C 232kPa 4:39  
136.2 C 231kPa 5:09  
136.2 C 231kPa 5:39  
136.2 C 232kPa 6:09  
136.2 C 231kPa 6:39  
136.2 C 231kPa 7:09  
136.2 C 231kPa 7:09  
VENTING 7:09  
AIR DRYING 7:59  
CYCLE COMPLETE 23:00  
-----

Digital Signature #  
4DCAEEA33E977ECA  
-----