
Midmark M11 - v1.0.0
Total Cycles: 2108

Name: RITTER 1
Sterilizer ID: V2178198
07/08/2022 03:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.4 F 0.0
2:00 190.4 F 2.9
4:00 210.9 F 8.7
6:00 223.0 F 8.4
8:00 232.7 F 9.2
10:00 239.1 F 10.0
12:00 242.6 F 12.1
14:00 241.0 F 10.7
16:00 254.2 F 17.5
18:00 265.8 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.1 F 28.5
2:00 272.3 F 29.0
3:00 271.7 F 28.3
4:00 272.3 F 29.0

Min 271.2 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.6 F 0.0
5:00 238.7 F 0.0
10:00 218.4 F 0.0
15:00 211.6 F 0.0
20:00 195.1 F 0.0
25:00 182.6 F 0.0
30:00 179.2 F 0.0
Drying Complete

FILLING: 2:31
HEATING: 19:36
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:18
CYCLE COMPLETE

07/08/2022 04:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____