
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002070
13:03 19/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 3:45
STERILIZING 4:46
136.4 C 224kPa 4:46
136.3 C 223kPa 5:16
136.4 C 224kPa 5:46
136.3 C 223kPa 6:16
136.3 C 223kPa 6:46
136.4 C 225kPa 7:16
136.4 C 224kPa 7:46
136.2 C 224kPa 8:16
136.2 C 224kPa 8:16
VENTING 8:16
AIR DRYING 9:11
CYCLE COMPLETE 9:15

Digital Signature #
AA08B0A6FAFE4BAB
