
Midmark M11 - v1.0.0
Total Cycles: 5038

Name: RITTER 1
Sterilizer ID: V2178198
10/30/2024 12:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.9 F 0.0
2:00 131.9 F 0.1
4:00 151.7 F 0.9
6:00 169.2 F 1.1
8:00 190.0 F 3.4
10:00 206.2 F 6.8
12:00 220.5 F 10.5
14:00 224.9 F 6.7
16:00 232.0 F 9.8
18:00 237.1 F 9.6
20:00 242.6 F 12.2
22:00 240.3 F 10.5
24:00 252.7 F 16.6
26:00 263.5 F 22.8
28:00 272.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.2
1:00 271.8 F 28.6
2:00 272.4 F 28.7
3:00 271.7 F 28.2
4:00 271.5 F 28.4

Min 271.2 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 211.4 F 0.0
10:00 225.4 F 0.0
15:00 228.8 F 0.0
20:00 228.6 F 0.0
25:00 229.5 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 28:11
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:06:46
CYCLE COMPLETE

10/30/2024 01:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____