
Midmark M11 - v1.0.0
Total Cycles: 1354

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2021 01:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.8 F 0.0
2:00 191.4 F 3.9
4:00 194.0 F 8.0
6:00 228.3 F 12.9
8:00 229.2 F 7.7
10:00 240.6 F 12.9
12:00 237.8 F 10.0
14:00 238.3 F 8.5
16:00 248.7 F 14.5
18:00 260.3 F 21.0
20:00 270.9 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 29.0
2:00 271.6 F 28.2
3:00 272.3 F 29.1
4:00 272.3 F 28.7

Min 271.3 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 232.5 F 0.0
10:00 232.1 F 0.0
15:00 226.8 F 0.0
20:00 225.5 F 0.0
25:00 229.7 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 20:45
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:54
CYCLE COMPLETE

11/10/2021 02:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____