

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1312

Name: RITTER 1  
Sterilizer ID: V2178198  
10/28/2021 08:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.9 F 0.0  
2:00 189.5 F 1.8  
4:00 198.2 F 6.2  
6:00 221.8 F 12.2  
8:00 228.6 F 7.6  
10:00 242.0 F 13.2  
12:00 238.4 F 10.7  
14:00 236.8 F 8.9  
16:00 250.8 F 16.0  
18:00 263.0 F 23.0  
20:00 271.7 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.5 F 29.0  
2:00 272.6 F 29.4  
3:00 272.1 F 28.7  
4:00 272.9 F 29.5

Min 271.6 F 28.5  
Max 273.3 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.7 F 0.0  
5:00 228.6 F 0.0  
10:00 224.6 F 0.0  
15:00 224.6 F 0.0  
20:00 225.2 F 0.0  
25:00 229.1 F 0.0  
30:00 229.6 F 0.0  
Drying Complete

FILLING: 1:56  
HEATING: 20:04  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:59:07  
CYCLE COMPLETE

10/28/2021 09:04 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_