
Midmark M11 - v1.0.0
Total Cycles: 5175

Name: RITTER 1
Sterilizer ID: V2178198
12/10/2024 03:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 87.1 F 0.0
2:00 92.1 F 0.0
4:00 123.2 F 0.0
6:00 149.9 F 0.1
8:00 173.6 F 0.2
10:00 189.8 F 0.3
12:00 203.0 F 0.7
14:00 217.8 F 3.2
16:00 225.6 F 6.1
18:00 233.5 F 9.5
20:00 242.9 F 13.4
22:00 239.8 F 10.5
24:00 237.0 F 8.7
26:00 244.9 F 12.0
28:00 245.5 F 12.7
30:00 256.2 F 18.3
32:00 265.4 F 23.9
34:00 272.2 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.0
1:00 271.6 F 27.9
2:00 271.5 F 27.9
3:00 271.6 F 28.0
4:00 271.7 F 28.1

Min 271.5 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 212.7 F 0.0
10:00 226.1 F 0.0
15:00 237.5 F 0.0
20:00 230.7 F 0.0
25:00 223.5 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 34:07
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:12:40
CYCLE COMPLETE

12/10/2024 04:37 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____