

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000250  
15:46 29/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:57  
PRESSURIZING 3:06  
STERILIZING 4:33  
136.0 C 230kPa 4:33  
136.2 C 232kPa 5:03  
136.4 C 233kPa 5:33  
136.4 C 233kPa 6:03  
136.2 C 232kPa 6:33  
136.4 C 233kPa 7:03  
136.4 C 233kPa 7:33  
136.3 C 232kPa 8:03  
136.3 C 232kPa 8:03  
VENTING 8:03  
AIR DRYING 9:07  
CYCLE COMPLETE 9:09  
-----

Digital Signature #  
9EFED60B0870EC42  
-----