

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000188  
13:40 17/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:30  
PRESSURIZING 2:33  
STERILIZING 3:37  
136.4 C 226kPa 3:37  
136.3 C 225kPa 4:07  
136.4 C 223kPa 4:37  
136.3 C 225kPa 5:07  
136.3 C 224kPa 5:37  
136.4 C 225kPa 6:07  
136.2 C 222kPa 6:37  
136.4 C 226kPa 7:07  
136.4 C 226kPa 7:07  
VENTING 7:07  
AIR DRYING 8:00  
CYCLE COMPLETE 8:45

-----  
Digital Signature #  
84F3F2F470199BCF  
-----