

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000263  
15:24 30/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:51  
PRESSURIZING 3:52  
STERILIZING 5:10  
136.0 C 222kPa 5:10  
136.4 C 233kPa 5:40  
136.0 C 231kPa 6:10  
136.4 C 233kPa 6:40  
136.4 C 233kPa 7:10  
136.3 C 231kPa 7:40  
136.3 C 233kPa 8:10  
136.1 C 232kPa 8:40  
136.1 C 232kPa 8:40  
VENTING 8:40  
AIR DRYING 9:40  
CYCLE COMPLETE 24:40  
-----

Digital Signature #  
78B39636C0F8E0DC  
-----