

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001747  
15:01 17/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:05  
PRESSURIZING 2:09  
STERILIZING 3:06  
136.4 C 220kPa 3:06  
136.3 C 225kPa 3:36  
136.3 C 225kPa 4:06  
136.3 C 225kPa 4:36  
136.2 C 225kPa 5:06  
136.2 C 224kPa 5:36  
136.4 C 227kPa 6:06  
136.4 C 227kPa 6:36  
136.4 C 227kPa 6:36  
VENTING 6:36  
AIR DRYING 7:30  
CYCLE COMPLETE 7:40  
-----

Digital Signature #  
58B81548E0B62292  
-----