
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000204
7:09 08/07/2020

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:25
STERILIZING 3:30
136.0 C 220kPa 3:30
136.2 C 220kPa 4:00
136.3 C 222kPa 4:30
136.3 C 223kPa 5:00
136.3 C 221kPa 5:30
136.2 C 222kPa 6:00
136.2 C 221kPa 6:30
136.4 C 223kPa 7:00
136.4 C 223kPa 7:00
VENTING 7:00
AIR DRYING 7:49
CYCLE COMPLETE 7:51

Digital Signature #
48373739368DB5CB
