
Midmark M11 - v1.0.5
Total Cycles: 3471

Name: RITTER2
Sterilizer ID: V2313228
03/22/2024 11:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.8 F 0.0
2:00 169.3 F 2.3
4:00 182.4 F 6.2
6:00 200.1 F 10.6
8:00 220.9 F 5.6
10:00 226.6 F 9.5
12:00 234.7 F 9.7
14:00 243.7 F 12.5
16:00 242.2 F 11.4
18:00 254.6 F 17.9
20:00 265.6 F 24.5
22:00 272.0 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.9 F 29.7
2:00 272.0 F 29.1
3:00 271.5 F 29.2
4:00 272.5 F 29.7

Min 271.3 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.2 F 0.0
5:00 234.4 F 0.0
10:00 204.5 F 0.0
15:00 200.2 F 0.0
20:00 197.0 F 0.0
25:00 194.7 F 0.0
30:00 192.7 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 22:37
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:01:28
CYCLE COMPLETE

03/22/2024 12:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____