
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 002687
11:54 15/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:19
PRESSURIZING 3:20
STERILIZING 4:23
136.0 C 220kPa 4:23
135.9 C 220kPa 4:53
136.1 C 222kPa 5:23
136.2 C 223kPa 5:53
136.4 C 222kPa 6:23
136.4 C 222kPa 6:53
136.4 C 222kPa 7:23
136.4 C 221kPa 7:53
136.4 C 221kPa 7:53
VENTING 7:53
AIR DRYING 8:49
CYCLE COMPLETE 9:29

Digital Signature #
88CECB9ACABE8C1F
