
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001568
12:38 07/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:07
STERILIZING 4:08
136.0 C 220kPa 4:08
136.3 C 224kPa 4:38
136.4 C 222kPa 5:08
136.3 C 225kPa 5:38
136.4 C 226kPa 6:08
136.4 C 225kPa 6:38
136.4 C 226kPa 7:08
136.3 C 224kPa 7:38
136.3 C 224kPa 7:38
VENTING 7:38
AIR DRYING 8:31
CYCLE COMPLETE 9:10

Digital Signature #
CE59CB05E0F8C088
