
Midmark M11 - v1.0.5
Total Cycles: 3098

Name: RITTER2
Sterilizer ID: V2313228
11/08/2023 10:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.9 F 0.0
2:00 159.8 F 1.3
4:00 167.7 F 4.5
6:00 190.0 F 8.8
8:00 205.9 F 13.5
10:00 220.5 F 8.0
12:00 233.7 F 12.8
14:00 235.5 F 10.1
16:00 236.6 F 8.8
18:00 248.2 F 14.9
20:00 259.6 F 21.3
22:00 270.1 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 273.0 F 29.9
2:00 272.1 F 29.3
3:00 271.1 F 29.0
4:00 272.0 F 29.9

Min 271.1 F 28.9
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 214.1 F 0.0
10:00 213.3 F 0.0
15:00 212.0 F 0.0
20:00 210.5 F 0.0
25:00 213.0 F 0.0
30:00 209.3 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 22:57
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:01:40
CYCLE COMPLETE

11/08/2023 11:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____