
Midmark M11 - v1.0.0
Total Cycles: 3116

Name: RITTER 1
Sterilizer ID: V2178198
04/14/2023 02:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.5 F 0.0
2:00 165.9 F 1.2
4:00 186.8 F 5.4
6:00 214.1 F 10.5
8:00 221.2 F 5.3
10:00 234.8 F 10.7
12:00 233.5 F 7.3
14:00 244.0 F 12.2
16:00 244.7 F 12.7
18:00 257.6 F 19.5
20:00 268.9 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.5 F 29.1
2:00 271.4 F 28.4
3:00 272.6 F 29.2
4:00 271.8 F 28.5

Min 271.3 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 226.9 F 0.0
10:00 226.6 F 0.0
15:00 224.7 F 0.0
20:00 219.4 F 0.0
25:00 218.3 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:06
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:36
CYCLE COMPLETE

04/14/2023 03:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____