
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001242
13:22 01/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:24
PRESSURIZING 3:26
STERILIZING 4:32
136.0 C 220kPa 4:32
136.4 C 223kPa 5:02
136.4 C 223kPa 5:32
136.3 C 224kPa 6:02
136.3 C 223kPa 6:32
136.4 C 224kPa 7:02
136.4 C 225kPa 7:32
136.4 C 223kPa 8:02
136.4 C 223kPa 8:02
VENTING 8:02
AIR DRYING 8:57
CYCLE COMPLETE 23:57

Digital Signature #
28EC898E490CA0F6
