

-----  
Midmark M11 - v1.0.0  
Total Cycles: 953

Name: RITTER 1  
Sterilizer ID: V2178198  
06/29/2021 10:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.6 F 0.0  
2:00 157.9 F 1.0  
4:00 161.1 F 2.7  
6:00 181.4 F 5.6  
8:00 196.7 F 8.8  
10:00 210.9 F 12.5  
12:00 220.1 F 5.2  
14:00 230.8 F 9.4  
16:00 238.8 F 11.4  
18:00 242.1 F 11.6  
20:00 240.3 F 10.4  
22:00 252.3 F 16.2  
24:00 262.6 F 22.2  
26:00 272.2 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.0  
1:00 272.5 F 28.8  
2:00 272.1 F 28.3  
3:00 271.4 F 28.0  
4:00 272.1 F 28.7

Min 271.4 F 27.9  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.7 F 0.0  
5:00 237.2 F 0.0  
10:00 231.1 F 0.0  
15:00 219.6 F 0.0  
20:00 212.6 F 0.0  
25:00 214.0 F 0.0  
30:00 199.6 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 26:34  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:05:21  
CYCLE COMPLETE

06/29/2021 11:54 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_