
Midmark M11 - v1.0.0
Total Cycles: 803

Name: RITTER 1
Sterilizer ID: V2178198
04/22/2021 03:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.9 F 0.0
2:00 189.1 F 3.8
4:00 201.5 F 9.7
6:00 225.4 F 6.1
8:00 232.7 F 12.1
10:00 235.3 F 9.7
12:00 235.8 F 8.3
14:00 250.6 F 15.8
16:00 263.5 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.3 F 28.9
2:00 271.7 F 29.2
3:00 272.7 F 29.2
4:00 272.9 F 29.6

Min 271.0 F 28.4
Max 273.4 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.8 F 0.0
5:00 233.5 F 0.0
10:00 220.8 F 0.0
15:00 216.7 F 0.0
20:00 213.7 F 0.0
25:00 213.3 F 0.0
Dry Cycle Aborted

FILLING: 2:06
HEATING: 17:55
STERILIZING: 4:00
VENTING: 3:03
DRYING: 27:31
TOTAL CYCLE: 00:54:34

04/26/2021 06:22 AM
ACCEPTED: X
REJECTED:
Operator: U02