

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5209

Name: RITTER 1  
Sterilizer ID: V2178198  
12/20/2024 09:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.9 F 0.0  
2:00 73.9 F 0.0  
4:00 89.7 F 0.0  
6:00 120.8 F 0.0  
8:00 150.0 F 0.1  
10:00 176.0 F 0.1  
12:00 200.2 F 0.5  
14:00 213.9 F 3.2  
16:00 223.2 F 6.0  
18:00 231.4 F 9.2  
20:00 240.4 F 13.0  
22:00 237.4 F 9.5  
24:00 239.8 F 9.7  
26:00 246.5 F 13.2  
28:00 243.7 F 11.7  
30:00 256.2 F 18.3  
32:00 267.3 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 27.9  
1:00 272.8 F 28.7  
2:00 271.9 F 27.9  
3:00 272.3 F 28.5  
4:00 272.6 F 28.4  
  
Min 271.6 F 27.8  
Max 272.9 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.1 F 0.0  
5:00 197.0 F 0.0  
10:00 189.5 F 0.0  
15:00 196.5 F 0.0  
20:00 218.9 F 0.0  
25:00 231.0 F 0.0  
30:00 224.8 F 0.0  
  
Drying Complete

FILLING: 0:41  
HEATING: 33:29  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:11:33  
CYCLE COMPLETE

12/20/2024 10:43 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_