
Midmark M11 - v1.0.5
Total Cycles: 1961

Name: RITTER2
Sterilizer ID: V2313228
01/23/2023 04:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.7 F 0.0
2:00 182.7 F 1.1
4:00 191.9 F 5.8
6:00 211.2 F 11.2
8:00 224.6 F 7.1
10:00 238.2 F 12.8
12:00 238.5 F 11.0
14:00 238.0 F 10.0
16:00 252.7 F 17.3
18:00 264.8 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.3 F 29.9
2:00 272.1 F 29.4
3:00 271.2 F 29.0
4:00 271.6 F 29.7

Min 271.1 F 28.9
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.6 F 0.0
5:00 202.1 F 0.0
10:00 217.5 F 0.0
15:00 232.5 F 0.0
20:00 229.5 F 0.0
25:00 235.0 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 19:53
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:36
CYCLE COMPLETE

01/23/2023 05:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____