
Midmark M11 - v1.0.0
Total Cycles: 3793

Name: RITTER 1
Sterilizer ID: V2178198
09/19/2023 09:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 75.7 F 0.0
4:00 84.9 F 0.0
6:00 120.9 F 0.0
8:00 151.0 F 1.1
10:00 178.6 F 3.0
12:00 198.0 F 5.3
14:00 213.8 F 8.8
16:00 224.3 F 12.3
18:00 227.0 F 6.3
20:00 237.1 F 10.5
22:00 233.8 F 8.0
24:00 243.4 F 12.2
26:00 239.9 F 10.3
28:00 252.7 F 16.5
30:00 263.5 F 22.8
32:00 272.3 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 271.9 F 28.4
2:00 272.6 F 28.8
3:00 271.6 F 28.5
4:00 272.0 F 28.3

Min 271.4 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.6 F 0.0
5:00 223.2 F 0.0
10:00 220.8 F 0.0
15:00 218.6 F 0.0
20:00 210.4 F 0.0
25:00 211.7 F 0.0
30:00 206.6 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 32:11
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:10:18
CYCLE COMPLETE

09/19/2023 11:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____