
Midmark M11 - v1.0.5
Total Cycles: 616

Name: RITTER2
Sterilizer ID: V2313228
12/08/2021 04:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.4 F 0.0
2:00 148.4 F 2.7
4:00 171.7 F 7.1
6:00 193.5 F 12.4
8:00 224.2 F 6.3
10:00 232.4 F 11.5
12:00 233.5 F 8.8
14:00 239.5 F 10.2
16:00 247.5 F 14.4
18:00 260.1 F 21.5
20:00 272.0 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.9 F 29.6
2:00 271.8 F 28.7
3:00 272.6 F 29.7
4:00 272.6 F 29.4

Min 271.6 F 28.7
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.3 F 0.0
5:00 229.1 F 0.0
10:00 230.1 F 0.0
15:00 225.0 F 0.0
20:00 225.0 F 0.0
25:00 229.5 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 20:35
STERILIZING: 4:00
VENTING: 2:54
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

12/09/2021 07:31 AM
ACCEPTED: X
REJECTED:
Operator: U02