
Midmark M11 - v1.0.0
Total Cycles: 3685

Name: RITTER 1
Sterilizer ID: V2178198
08/25/2023 02:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.3 F 0.0
2:00 165.2 F 2.6
4:00 173.8 F 5.7
6:00 196.8 F 10.1
8:00 223.5 F 6.1
10:00 229.9 F 8.2
12:00 239.4 F 12.7
14:00 236.7 F 9.4
16:00 239.4 F 10.0
18:00 246.1 F 13.3
20:00 257.3 F 19.3
22:00 267.7 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.2
1:00 272.9 F 29.2
2:00 272.0 F 29.0
3:00 272.0 F 28.6
4:00 272.8 F 29.1

Min 270.9 F 28.1
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.2 F 0.0
5:00 221.9 F 0.0
10:00 228.5 F 0.0
15:00 190.7 F 0.0
20:00 191.5 F 0.0
25:00 198.7 F 0.0
30:00 212.7 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 23:22
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:02:18
CYCLE COMPLETE

08/25/2023 03:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____