
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002561
14:02 05/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:44
STERILIZING 3:47
136.2 C 221kPa 3:47
136.4 C 224kPa 4:17
136.3 C 224kPa 4:47
136.4 C 225kPa 5:17
136.3 C 224kPa 5:47
136.4 C 224kPa 6:17
136.3 C 223kPa 6:47
136.4 C 225kPa 7:17
136.4 C 225kPa 7:17
VENTING 7:18
AIR DRYING 8:08
CYCLE COMPLETE 8:11

Digital Signature #
F0B20F2CB2AAF625
