

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001488  
8:56 10/08/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:27  
PRESSURIZING 7:20  
STERILIZING 8:58  
136.0 C 220kPa 8:58  
136.2 C 221kPa 9:28  
136.3 C 221kPa 9:59  
136.1 C 218kPa 10:29  
136.4 C 217kPa 10:59  
136.3 C 222kPa 11:29  
136.3 C 221kPa 11:59  
135.9 C 220kPa 12:28  
VENTING 12:29  
AIR DRYING 13:34  
CYCLE COMPLETE 25:20  
-----

Digital Signature #  
25734306989F3B95  
-----