

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002139  
11:08 10/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:04  
PRESSURIZING 2:08  
STERILIZING 3:16  
136.4 C 224kPa 3:16  
136.4 C 225kPa 3:46  
136.4 C 225kPa 4:16  
136.2 C 221kPa 4:46  
136.4 C 222kPa 5:16  
136.3 C 224kPa 5:46  
136.4 C 225kPa 6:16  
136.3 C 224kPa 6:46  
136.3 C 224kPa 6:46  
VENTING 6:46  
AIR DRYING 7:36  
CYCLE COMPLETE 37:36  
-----

Digital Signature #  
75E0236E60B9CF91  
-----