
Midmark M11 - v1.0.0
Total Cycles: 4655

Name: RITTER 1
Sterilizer ID: V2178198
06/18/2024 04:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.0 F 0.0
2:00 116.2 F 1.3
4:00 137.0 F 4.0
6:00 169.1 F 7.5
8:00 193.4 F 11.8
10:00 208.3 F 4.1
12:00 225.9 F 8.2
14:00 236.6 F 12.7
16:00 233.5 F 8.7
18:00 243.0 F 11.4
20:00 243.5 F 12.1
22:00 255.5 F 18.3
24:00 266.4 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.8 F 29.1
2:00 271.7 F 28.3
3:00 272.5 F 29.2
4:00 272.5 F 28.8

Min 271.5 F 28.3
Max 273.1 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 210.5 F 0.0
10:00 222.1 F 0.0
15:00 227.9 F 0.0
20:00 223.8 F 0.0
25:00 224.2 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:39
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:04:03
CYCLE COMPLETE

06/18/2024 05:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____