

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 006462  
13:31 13/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:39  
PRESSURIZING 2:43  
STERILIZING 3:51  
136.5 C 222kPa 3:51  
136.2 C 224kPa 4:21  
136.2 C 221kPa 4:51  
136.4 C 223kPa 5:21  
136.4 C 225kPa 5:51  
136.4 C 224kPa 6:21  
136.2 C 222kPa 6:51  
136.0 C 221kPa 7:21  
136.0 C 221kPa 7:21  
VENTING 7:21  
AIR DRYING 8:11  
CYCLE COMPLETE 8:24  
-----

Digital Signature #  
8EB9C2CCE6FC520B  
-----