
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006117
11:43 15/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:28
PRESSURIZING 1:33
STERILIZING 2:30
136.0 C 220kPa 2:30
136.1 C 223kPa 3:00
136.2 C 224kPa 3:30
136.3 C 222kPa 4:00
136.4 C 225kPa 4:30
136.3 C 224kPa 5:00
136.2 C 222kPa 5:30
136.4 C 224kPa 6:00
136.4 C 224kPa 6:00
VENTING 6:00
AIR DRYING 6:52
CYCLE COMPLETE 7:07

Digital Signature #
F9D4CC52E2C0384A
