
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002098
10:30 01/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:01
STERILIZING 2:58
136.0 C 220kPa 2:58
136.1 C 223kPa 3:28
136.4 C 226kPa 3:58
136.4 C 225kPa 4:28
136.4 C 224kPa 4:58
136.3 C 224kPa 5:28
136.4 C 225kPa 5:58
136.4 C 225kPa 6:28
136.4 C 225kPa 6:28
VENTING 6:28
AIR DRYING 7:18
CYCLE COMPLETE 29:45

Digital Signature #
71574C68AB88F317
