
Midmark M11 - v1.0.0
Total Cycles: 2190

Name: RITTER 1
Sterilizer ID: V2178198
08/03/2022 07:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.6 F 0.0
2:00 73.0 F 0.0
4:00 77.3 F 0.0
6:00 110.2 F 0.2
8:00 142.2 F 2.3
10:00 168.6 F 5.2
12:00 186.1 F 8.9
14:00 201.1 F 12.6
16:00 213.3 F 4.6
18:00 225.3 F 8.3
20:00 235.7 F 12.4
22:00 234.5 F 8.7
24:00 242.7 F 11.5
26:00 244.0 F 12.3
28:00 255.7 F 18.4
30:00 266.2 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.0 F 28.9
2:00 272.1 F 28.6
3:00 271.6 F 28.6
4:00 272.5 F 29.0

Min 271.3 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.9 F 0.0
5:00 211.9 F 0.0
10:00 217.4 F 0.0
15:00 225.4 F 0.0
20:00 236.9 F 0.0
25:00 231.5 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 31:35
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:29
CYCLE COMPLETE

08/03/2022 08:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____